

## Deutsche Akkreditierungsstelle GmbH

Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV

Signatory to the Multilateral Agreements of EA, ILAC and IAF for Mutual Recognition

# Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

**Ahlborn Mess- und Regelungstechnik GmbH**  
**Eichenfeldstraße 1-3, 83607 Holzkirchen**

is competent under the terms of DIN EN ISO/IEC 17025:2018 to carry out calibrations in the following fields:

### Thermodynamic quantities

#### Temperature quantities

- Resistance thermometers
- Thermocouples
- Temperature indicators and simulators

#### Humidity quantities

- Devices for relative humidity

### Fluid quantities

- Velocity of gases

### Electrical quantities

#### DC and low frequency quantities

- DC voltage
- DC current
- DC resistance
- AC voltage
- AC current

### Time and frequency

- Frequency

The accreditation certificate shall only apply in connection with the notice of accreditation of 20.05.2019 with the accreditation number D-K-19342-01. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 7 pages.

Registration number of the certificate: **D-K -19342-01-00**

Braunschweig,  
26.05.2019

Dr. Heike Manke  
Head of Division

Translation issued:  
24.10.2019

  
Head of Division

*The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.*  
<https://www.dakks.de/en/content/accredited-bodies-dakks>

This document is a translation. The definitive version is the original German accreditation certificate.

See notes overleaf.

# Deutsche Akkreditierungsstelle GmbH

Office Berlin  
Spittelmarkt 10  
10117 Berlin

Office Frankfurt am Main  
Europa-Allee 52  
60327 Frankfurt am Main

Office Braunschweig  
Bundesallee 100  
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkkS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkkS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkkS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

## Deutsche Akkreditierungsstelle GmbH

### Annex to the Accreditation Certificate D-K-19342-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 20.05.2019

Date of issue: 20.05.2019

Holder of certificate:

**Ahlborn Mess- und Regelungstechnik GmbH**  
**Eichenfeldstraße 1-3, 83607 Holzkirchen**

Calibration in the fields:

#### Thermodynamic quantities

##### Temperature quantities

- Resistance thermometers
- Thermocouples
- Temperature indicators and simulators

##### Humidity quantities

- Devices for relative humidity

#### Fluid quantities

- Velocity of gases

#### Electrical quantities

##### DC and low frequency quantities

- DC voltage
- DC current
- DC resistance
- AC voltage
- AC current

#### Time and frequency

- Frequency

Abbreviations used: see last page

Within the measurands/calibration items marked with <sup>\*</sup>, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

**Annex to the accreditation certificate D-K-19342-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Relative humidity</b> Electrical hygrometers	10 % to 30 %	in 2-pressure-generators temperature range 20 °C to 30 °C AA070001:2019-01	0,4 %	Measurement uncertainty expressed in relative humidity
	> 30 % to 60 %		0,8 %	
	> 60 % to 85 %		1,0 %	
<b>Mechanical hygrometers</b>	10 % to 30 %	in climatic chambers air temperature: 10 °C to 95 °C AA070001:2019-01	0,7 %	Measurement uncertainty expressed in relative humidity
	> 30 % to 60 %		0,9 %	
	> 60 % to 95 %		1,1 %	
<b>Electrical hygrometers, psychrometers</b>	10 % to 98 %	in climatic chambers air temperature: 10 °C to 95 °C AA070001:2019-01	0,6 %	Comparison with standard psychrometer
<b>Temperature</b> Temperature measuring chains or temperature probes with resistance sensors*)	0,01 °C	triple point of water	2 mK	calibration at fixed point temperature
	0 °C	in ice baths DKD-R 5-1:2018	10 mK	Comparison with standard resistance thermometer
	20 °C to 30 °C	in 2-pressure-generators DKD-R 5-1:2018	0,06 K	
	-30 °C to < 0 °C	in climatic chambers DKD-R 5-1:2018	0,10 K	
	0 °C to 100 °C		0,08 K	
	> 100 °C to 150 °C		0,10 K	
	-100 °C to 200 °C	in liquid baths DKD-R 5-1:2018	10 mK	
	250 °C to 550 °C	in salt baths DKD-R 5-1:2018	20 mK	
	-100 °C to < -50 °C	in dry block calibrators DKD-R 5-1:2018	0,13 K	
	-50 °C to 250 °C		0,10 K	
	> 250 °C to 500 °C		0,15 K	
	> 500 °C to 650 °C		0,17 K	
	-100 °C to 155 °C	in dry block calibrators DKD-R 5-1:2018	0,03 K	Comparison with standard resistance thermometer
	> 155 °C to 250 °C		0,06 K	Only for calibration items with sensors in metal protection tubes
	> 250 °C to 650 °C		0,12 K	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-19342-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Temperature measuring chains or temperature probes with thermocouple sensors <sup>*)</sup>	20 °C to 30 °C	in 2-pressure-generators DKD-R 5-3:2018	0,3 K	Comparison with standard resistance thermometer
	-30 °C to 150 °C	in climatic chambers DKD-R 5-3:2018	0,3 K	
	-100 °C to 250 °C	in liquid baths or dry block calibrators DKD-R 5-3:2018	0,3 K	
	> 250 °C to 650 °C		0,8 K	
	500 °C to 1100 °C	in tube furnaces DKD-R 5-3:2018	1,5 K	Comparison with standard thermocouple
	> 1100 °C to 1200 °C		2,0 K	
Temperature indicators and simulators for resistance thermometers <sup>*)</sup>	-200 °C to 661 °C	DKD-R 5-5:2018	10 mK	Comparison with reference multimeter Characteristic curve according to DIN EN 60751:2009
	> 661 °C to 850 °C		20 mK	
Temperature indicators and simulators for base metal thermocouples <sup>*)</sup>	-200 °C to 1300 °C	DKD-R 5-5:2018	0,1 K	without reference junction compensation Characteristic curve according to DIN EN 60584-01:2014
	-200 °C to 1300 °C	DKD-R 5-5:2018	0,3 K	with reference junction compensation Characteristic curve according to DIN EN 60584-01:2014
Temperature indicators and simulators for noble metal thermocouples Type S, R <sup>*)</sup>	0 °C to 1768 °C	DKD-R 5-5:2010	0,2 K	without reference junction compensation Characteristic curve according to DIN EN 60584-01:2014
Temperature indicators and simulators for noble metal thermocouples Typ S, R <sup>*)</sup>	0 °C to 1768 °C	DKD-R 5-5:2010	0,3 K	with reference junction compensation Characteristic curve according to DIN EN 60584-01:2014

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-19342-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Temperature indicators and simulators for noble metal thermocouples Typ B <sup>*)</sup>	500 °C to 1820 °C	DKD-R 5-5:2010	0,2 K	without reference junction compensation Characteristic curve according to DIN EN 60584-01:2014
Temperature indicators and simulators for noble metal thermocouples Typ B <sup>*)</sup>	500 °C to 1820 °C	DKD-R 5-5:2010	0,3 K	with reference junction compensation Characteristic curve according to DIN EN 60584-01:2014
Velocity of gases	0,1 m/s to 65 m/s	AA070002: 2019-01	0,5 %, but not smaller than 0,01 m/s	wind tunnel: type Göttingen nozzle: 320 mm  calibration standard: laser-Doppler anemometer
DC voltage Sources	0 V to < 0,2 V		$0,2 \mu\text{V} + 5 \cdot 10^{-6} \cdot U$	U: measured value
	0,2 V to < 2 V		$0,5 \mu\text{V} + 5 \cdot 10^{-6} \cdot U$	
	2 V to < 20 V		$5 \mu\text{V} + 5 \cdot 10^{-6} \cdot U$	
	20 V to < 200 V		$50 \mu\text{V} + 5 \cdot 10^{-6} \cdot U$	
	200 V to 1000 V		$0,5 \text{ mV} + 5 \cdot 10^{-6} \cdot U$	
Measuring instruments	0 V to < 0,33 V		$0,8 \mu\text{V} + 22 \cdot 10^{-6} \cdot U$	U: measured value
	0,33 V to < 3,3 V		$1,4 \mu\text{V} + 10 \cdot 10^{-6} \cdot U$	
	3,3 V to < 33 V		$16 \cdot 10^{-6} \cdot U$	
	33 V to < 330 V		$0,1 \text{ mV} + 16 \cdot 10^{-6} \cdot U$	
	330 V to 1020 V		$1,0 \text{ mV} + 16 \cdot 10^{-6} \cdot U$	
DC current Sources	1 $\mu\text{A}$ to < 200 $\mu\text{A}$		$0,4 \text{ nA} + 15 \cdot 10^{-6} \cdot I$	I: measured value
	200 $\mu\text{A}$ to < 2 mA		$4,0 \text{ nA} + 15 \cdot 10^{-6} \cdot I$	
	2 mA to < 20 mA		$40 \text{ nA} + 15 \cdot 10^{-6} \cdot I$	
	20 mA to < 200 mA		$0,8 \mu\text{A} + 45 \cdot 10^{-6} \cdot I$	
	200 mA to < 2 A		$16 \mu\text{A} + 0,17 \cdot 10^{-3} \cdot I$	
	2 A to 20 A		$0,4 \text{ mA} + 0,37 \cdot 10^{-3} \cdot I$	
Measuring instruments	1 $\mu\text{A}$ to < 330 $\mu\text{A}$		$20 \text{ nA} + 0,1 \cdot 10^{-3} \cdot I$	I: measured value
	330 $\mu\text{A}$ to < 3,3 mA		$40 \text{ nA} + 77 \cdot 10^{-6} \cdot I$	
	3,3 mA to < 33 mA		$0,22 \mu\text{A} + 77 \cdot 10^{-6} \cdot I$	
	33 mA to < 330 mA		$2,2 \mu\text{A} + 77 \cdot 10^{-6} \cdot I$	
	330 mA to < 1,1 A		$30 \mu\text{A} + 0,16 \cdot 10^{-3} \cdot I$	
	1,1 A to < 3 A		$30 \mu\text{A} + 0,28 \cdot 10^{-3} \cdot I$	
	3 A to < 11 A		$0,3 \text{ mA} + 0,40 \cdot 10^{-3} \cdot I$	
	11 A to 20,5 A		$0,3 \text{ mA} + 0,78 \cdot 10^{-3} \cdot I$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-19342-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
DC resistance Sources	0,1 Ω to < 2 Ω		$4 \mu\Omega + 15 \cdot 10^{-6} \cdot R$	R: measured value
	2 Ω to < 20 Ω		$20 \mu\Omega + 9 \cdot 10^{-6} \cdot R$	
	20 Ω to < 200 Ω		$40 \mu\Omega + 8 \cdot 10^{-6} \cdot R$	
	200 Ω to < 2 kΩ		$10 \cdot 10^{-6} \cdot R$	
	2 kΩ to < 20 kΩ		$10 \cdot 10^{-6} \cdot R$	
	20 kΩ to < 200 kΩ		$10 \cdot 10^{-6} \cdot R$	
	200 kΩ to < 2 MΩ		$14 \cdot 10^{-6} \cdot R$	
	2 MΩ to < 20 MΩ		$0,1 \text{ k}\Omega + 20 \cdot 10^{-6} \cdot R$	
	20 MΩ to < 200 MΩ		$10 \text{ k}\Omega + 0,11 \cdot 10^{-3} \cdot R$	
	200 MΩ to 2 GΩ		$1 \text{ M}\Omega + 0,55 \cdot 10^{-3} \cdot R$	
Measuring instruments	0 Ω to < 11 Ω		$0,75 \text{ m}\Omega + 36 \cdot 10^{-6} \cdot R$	R: measured value
	11 Ω to < 33 Ω		$1,1 \text{ m}\Omega + 24 \cdot 10^{-6} \cdot R$	
	33 Ω to < 110 Ω		$1,0 \text{ m}\Omega + 22 \cdot 10^{-6} \cdot R$	
	110 Ω to < 330 Ω		$1,4 \text{ m}\Omega + 22 \cdot 10^{-6} \cdot R$	
	330 Ω to < 1,1 kΩ		$30 \cdot 10^{-6} \cdot R$	
	1,1 kΩ to < 3,3 kΩ		$35 \cdot 10^{-6} \cdot R$	
	3,3 kΩ to < 11 kΩ		$30 \cdot 10^{-6} \cdot R$	
	11 kΩ to < 33 kΩ		$35 \cdot 10^{-6} \cdot R$	
	33 kΩ to < 110 kΩ		$30 \cdot 10^{-6} \cdot R$	
	110 kΩ to < 330 kΩ		$40 \cdot 10^{-6} \cdot R$	
	330 kΩ to < 1,1 MΩ		$30 \cdot 10^{-6} \cdot R$	
	1,1 MΩ to < 3,3 MΩ		$80 \cdot 10^{-6} \cdot R$	
	3,3 MΩ to < 11 MΩ		$0,16 \cdot 10^{-3} \cdot R$	
	11 MΩ to < 33 MΩ		$0,30 \cdot 10^{-3} \cdot R$	
	33 MΩ to < 110 MΩ		$0,60 \cdot 10^{-3} \cdot R$	
	110 MΩ to < 330 MΩ		$3,5 \cdot 10^{-3} \cdot R$	
	330 MΩ to 1,1 GΩ		$18 \cdot 10^{-3} \cdot R$	
AC voltage Sources	10 mV to < 200 mV	10 Hz to 40 Hz	$4 \mu\text{V} + 0,14 \cdot 10^{-3} \cdot U$	U: measured value
		> 40 Hz to 10 kHz	$4 \mu\text{V} + 0,14 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 30 kHz	$25 \mu\text{V} + 0,25 \cdot 10^{-3} \cdot U$	
		> 30 kHz to 100 kHz	$0,1 \text{ mV} + 0,30 \cdot 10^{-3} \cdot U$	
	0,2 V to < 2 V	10 Hz to 40 Hz	$20 \mu\text{V} + 0,11 \cdot 10^{-3} \cdot U$	
		> 40 Hz to 10 kHz	$20 \mu\text{V} + 0,10 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 30 kHz	$40 \mu\text{V} + 0,20 \cdot 10^{-3} \cdot U$	
		> 30 kHz to 100 kHz	$0,2 \text{ mV} + 0,5 \cdot 10^{-3} \cdot U$	
	2 V to < 20 V	10 Hz to 40 Hz	$0,2 \text{ mV} + 0,11 \cdot 10^{-3} \cdot U$	
		> 40 Hz to 10 kHz	$0,2 \text{ mV} + 0,10 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 30 kHz	$0,4 \text{ mV} + 0,20 \cdot 10^{-3} \cdot U$	
		> 30 kHz to 100 kHz	$2 \text{ mV} + 0,50 \cdot 10^{-3} \cdot U$	
	20 V to < 200 V	10 Hz to 40 Hz	$2 \text{ mV} + 0,11 \cdot 10^{-3} \cdot U$	
		> 40 Hz to 10 kHz	$2 \text{ mV} + 90 \cdot 10^{-6} \cdot U$	
		> 10 kHz to 30 kHz	$4 \text{ mV} + 0,20 \cdot 10^{-3} \cdot U$	
		> 30 kHz to 100 kHz	$20 \text{ mV} + 0,50 \cdot 10^{-3} \cdot U$	
200 V to 1050 V	40 Hz to 10 kHz	$20 \text{ mV} + 0,12 \cdot 10^{-3} \cdot U$		

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-19342-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>AC voltage</b> Measuring instruments	1 mV to < 33 mV	10 Hz to 45 Hz	$5 \mu\text{V} + 0,6 \cdot 10^{-3} \cdot U$	U: measured value
		> 45 Hz to 10 kHz	$5 \mu\text{V} + 0,12 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 20 kHz	$5 \mu\text{V} + 0,15 \cdot 10^{-3} \cdot U$	
		> 20 kHz to 50 kHz	$5 \mu\text{V} + 0,75 \cdot 10^{-3} \cdot U$	
		> 50 kHz to 100 kHz	$10 \mu\text{V} + 3 \cdot 10^{-3} \cdot U$	
	33 mV to < 330 mV	10 Hz to 45 Hz	$5 \mu\text{V} + 0,40 \cdot 10^{-3} \cdot U$	
		> 45 Hz to 10 kHz	$5 \mu\text{V} + 0,18 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 20 kHz	$5 \mu\text{V} + 0,25 \cdot 10^{-3} \cdot U$	
		> 20 kHz to 50 kHz	$5 \mu\text{V} + 0,50 \cdot 10^{-3} \cdot U$	
		> 50 kHz to 100 kHz	$20 \mu\text{V} + 0,80 \cdot 10^{-3} \cdot U$	
	330 mV to < 3,3 V	10 Hz to 45 Hz	$35 \mu\text{V} + 0,28 \cdot 10^{-3} \cdot U$	
		> 45 Hz to 10 kHz	$25 \mu\text{V} + 0,18 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 20 kHz	$45 \mu\text{V} + 0,20 \cdot 10^{-3} \cdot U$	
		> 20 kHz to 50 kHz	$25 \mu\text{V} + 0,30 \cdot 10^{-3} \cdot U$	
		> 50 kHz to 100 kHz	$0,80 \cdot 10^{-3} \cdot U$	
	3,3 V to < 33 V	10 Hz to 45 Hz	$0,30 \text{ mV} + 0,30 \cdot 10^{-3} \cdot U$	
		> 45 Hz to 10 kHz	$0,30 \text{ mV} + 0,20 \cdot 10^{-3} \cdot U$	
		> 10 kHz to 20 kHz	$0,15 \text{ mV} + 0,28 \cdot 10^{-3} \cdot U$	
		> 20 kHz to 50 kHz	$0,15 \text{ mV} + 0,35 \cdot 10^{-3} \cdot U$	
		> 50 kHz to 100 kHz	$0,15 \text{ mV} + 1,0 \cdot 10^{-3} \cdot U$	
33 V to < 330 V	45 Hz to 1 kHz	$0,2 \text{ mV} + 0,20 \cdot 10^{-3} \cdot U$		
	> 1 kHz to 10 kHz	$0,2 \text{ mV} + 0,28 \cdot 10^{-3} \cdot U$		
	> 10 kHz to 20 kHz	$0,1 \text{ mV} + 0,35 \cdot 10^{-3} \cdot U$		
	> 20 kHz to 50 kHz	$4 \text{ mV} + 0,50 \cdot 10^{-3} \cdot U$		
	> 50 kHz to 100 kHz	$30 \text{ mV} + 2,0 \cdot 10^{-3} \cdot U$		
330 V to 1020 V	45 Hz to 10 kHz	$0,25 \cdot 10^{-3} \cdot U$		
<b>AC current</b> Sources	1 $\mu\text{A}$ to < 200 $\mu\text{A}$	1 Hz to 10 kHz	$0,03 \mu\text{A} + 0,43 \cdot 10^{-3} \cdot I$	I: measured value
	200 $\mu\text{A}$ to < 2 mA	10 Hz to 10 kHz	$0,30 \mu\text{A} + 0,28 \cdot 10^{-3} \cdot I$	
	2 mA to < 20 mA	10 Hz to 10 kHz	$2,2 \mu\text{A} + 0,32 \cdot 10^{-3} \cdot I$	
	20 mA to < 200 mA	10 Hz to 10 kHz	$22 \mu\text{A} + 0,3 \cdot 10^{-3} \cdot I$	
	200 mA to < 2 A	10 Hz to 2 kHz	$0,22 \text{ mA} + 0,65 \cdot 10^{-3} \cdot I$	
	2 A to 20 A	10 Hz to 2 kHz	$2 \text{ mA} + 0,85 \cdot 10^{-3} \cdot I$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.



Annex to the accreditation certificate D-K-19342-01-00

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>AC current</b> Measuring instruments	29 µA to < 330 µA	45 Hz to 1 kHz	$0,1 \mu\text{A} + 1 \cdot 10^{-3} \cdot f$	f: measured value
		> 1 kHz to 5 kHz	$0,2 \mu\text{A} + 3 \cdot 10^{-3} \cdot f$	
		> 5 kHz to 10 kHz	$0,2 \mu\text{A} + 6 \cdot 10^{-3} \cdot f$	
	330 µA to < 3,3 mA	45 Hz to 1 kHz	$0,15 \mu\text{A} + 0,8 \cdot 10^{-3} \cdot f$	
		> 1 kHz to 5 kHz	$0,30 \mu\text{A} + 2 \cdot 10^{-3} \cdot f$	
		> 5 kHz to 10 kHz	$0,30 \mu\text{A} + 4 \cdot 10^{-3} \cdot f$	
	3,3 mA to < 33 mA	45 Hz to 1 kHz	$2 \mu\text{A} + 0,5 \cdot 10^{-3} \cdot f$	
		> 1 kHz to 5 kHz	$3 \mu\text{A} + 0,7 \cdot 10^{-3} \cdot f$	
		> 5 kHz to 10 kHz	$3 \mu\text{A} + 1,5 \cdot 10^{-3} \cdot f$	
	33 mA to < 330 mA	45 Hz to 1 kHz	$20 \mu\text{A} + 0,4 \cdot 10^{-3} \cdot f$	
		> 1 kHz to 5 kHz	$0,1 \text{ mA} + 0,8 \cdot 10^{-3} \cdot f$	
		> 5 kHz to 10 kHz	$0,1 \text{ mA} + 1,6 \cdot 10^{-3} \cdot f$	
	330 mA to < 1,1 A	45 Hz to 1 kHz	$50 \mu\text{A} + 0,45 \cdot 10^{-3} \cdot f$	
		> 1 kHz to 5 kHz	$0,7 \text{ mA} + 0,6 \cdot 10^{-3} \cdot f$	
	1,1 A to < 3 A	45 Hz to 1 kHz	$0,1 \text{ mA} + 0,6 \cdot 10^{-3} \cdot f$	
> 1 kHz to 5 kHz		$0,7 \text{ mA} + 5 \cdot 10^{-3} \cdot f$		
3 A to < 11 A	45 Hz to 100 Hz	$2 \text{ mA} + 0,6 \cdot 10^{-3} \cdot f$		
	> 100 Hz to 1 kHz	$2 \text{ mA} + 1 \cdot 10^{-3} \cdot f$		
11 A to 20,5 A	45 Hz to 100 Hz	$5 \text{ mA} + 1,2 \cdot 10^{-3} \cdot f$		
	> 100 Hz to 1 kHz	$5 \text{ mA} + 1,5 \cdot 10^{-3} \cdot f$		
<b>Frequency Sources</b>	10 Hz to 1,0 MHz		$20 \mu\text{Hz} + 13 \cdot 10^{-6} \cdot f$	f: measured value
Measuring instruments	0,01 Hz to < 120 Hz		$11 \text{ mHz} + 1 \cdot 10^{-6} \cdot f$	f: measured value
	120 Hz to < 1,2 kHz		$0,11 \text{ Hz} + 1 \cdot 10^{-6} \cdot f$	
	1,2 kHz to < 12 kHz		$1,1 \text{ Hz} + 1 \cdot 10^{-6} \cdot f$	
	12 kHz to < 0,12 MHz		$11 \text{ Hz} + 1 \cdot 10^{-6} \cdot f$	
	0,12 MHz to < 1,2 MHz		$0,11 \text{ kHz} + 1 \cdot 10^{-6} \cdot f$	
	1,2 MHz to 2,0 MHz		$1,1 \text{ kHz} + 1 \cdot 10^{-6} \cdot f$	

**Abbreviations used:**

AA	In house procedure of Ahlborn Mess- und Regelungstechnik GmbH
CMC	Calibration and measurement capabilities
DKD-R	Calibration Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt (PTB)

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.